# e-CHECK* (XS) Quality Control for the **ProCyte Dx*** Analyzer
# e-CHECK* (XS) contrôle qualité pour l'analyseur **ProCyte Dx***
# e-CHECK* (XS) Control de calidad para el analizador **ProCyte Dx***
# e-CHECK* (XS) Qualitätskontrolle für das **ProCyte Dx*** Analysegerät
# e-CHECK* (XS) Controllo qualita' per l'analizzatore **ProCyte Dx***
# e-CHECK* (XS) IDEXX**プロサイトDx***の精度管理

## L1 (Low / Bas / Bajo / Niedrig / Basso / 低濃度)

LOT

| Parameter<br>Paramètre<br>Parámetro<br>Parameter<br>Parametro<br>項目 | Target<br>Valeur cible<br>Objetivo<br>Ergebnis<br>Valore ideale<br>目標値 | Expected Range<br>Intervalle attendu<br>Rango esperado<br>Referenzbereich<br>Intervallo atteso<br>基準範囲 |
|---|---|---|
| RBC/GR ($10^6/\mu L$) | | |
| HCT (%) | | |
| HGB (g/dL) | | |
| MCV/VGM (fL) | | |
| MCH/TCMH (pg) | | |
| MCHC/CCMH (g/dL) | | |
| PLT ($K/\mu L$) | | |
| RDW-SD (fL) | | |
| RDW-CV (%) | | |
| PDW (fL) | | |
| PCT (%) | | |
| MPV (fL) | | |
| P-LCR (%) | | |
| WBC/GB ($K/\mu L$) | | |
| NEUT# ($K/\mu L$) | | |
| LYMPH# ($K/\mu L$) | | |
| MONO# ($K^9/\mu L$) | | |
| EO# ($K/\mu L$) | | |
| RBC-O ($10^6/\mu L$) | | |
| PLT-O ($K/\mu L$) | | |
| RETIC ($K/\mu L$) | | |
| % RETIC (%) | | |

## L2 (Normal / Normale / 正常濃度)

LOT 00280805 04/19/20

| Parameter<br>Paramètre<br>Parámetro<br>Parameter<br>Parametro<br>項目 | Target<br>Valeur cible<br>Objetivo<br>Ergebnis<br>Valore ideale<br>目標値 | Expected Range<br>Intervalle attendu<br>Rango esperado<br>Referenzbereich<br>Intervallo atteso<br>基準範囲 |
|---|---|---|
| RBC/GR ($10^6/\mu L$) | 4.27 | 4.00 - 4.54 |
| HCT (%) | 35.5 | 32.0 - 39.0 |
| HGB (g/dL) | 12.1 | 11.6 - 12.6 |
| MCV/VGM (fL) | 83.2 | 75.2 - 91.2 |
| MCH/TCMH (pg) | 28.3 | 24.6 - 32.0 |
| MCHC/CCMH (g/dL) | 34.1 | 28.0 - 40.2 |
| PLT ($K/\mu L$) | 228 | 183 - 273 |
| RDW-SD (fL) | 43.0 | 36.0 - 50.0 |
| RDW-CV (%) | 15.4 | 12.4 - 18.4 |
| PDW (fL) | 8.5 | 5.5 - 11.5 |
| PCT (%) | 0.20 | 0.12 - 0.28 |
| MPV (fL) | 8.9 | 5.9 - 11.9 |
| P-LCR (%) | 10.2 | 4.3 - 16.1 |
| WBC/GB ($K/\mu L$) | 7.04 | 6.29 - 7.79 |
| NEUT# ($K/\mu L$) | 3.35 | 2.29 - 4.41 |
| LYMPH# ($K/\mu L$) | 2.17 | 1.55 - 2.79 |
| MONO# ($K^9/\mu L$) | 0.83 | 0.07 - 1.59 |
| EO# ($K/\mu L$) | 0.69 | 0.33 - 1.05 |
| RBC-O ($10^6/\mu L$) | 4.14 | 3.44 - 4.84 |
| PLT-O ($K/\mu L$) | 207 | 157 - 257 |
| RETIC ($K/\mu L$) | 105.7 | 75.7 - 135.7 |
| % RETIC (%) | 2.48 | 1.68 - 3.28 |

Patent information: idexx.com/patents.

© 2013 IDEXX Laboratories, Inc. All rights reserved. • 06-0002107-00
*e-CHECK is a trademark of Sysmex Corporation. ProCyte Dx is a trademark or registered trademark of IDEXX Laboratories, Inc. or its affiliates in the United States and/or other countries.

IDEXX BioResearch