

- e-CHECK* (XS)** Quality Control for the **ProCyte Dx*** Analyzer
- e-CHECK* (XS)** contrôle qualité pour l'analyseur **ProCyte Dx***
- e-CHECK* (XS)** Control de calidad para el analizador **ProCyte Dx***
- e-CHECK* (XS)** Qualitätskontrolle für das **ProCyte Dx*** Analysegerät
- e-CHECK* (XS)** Controllo qualita' per l'analizzatore **ProCyte Dx***
- e-CHECK* (XS)** IDEXXプロサイトDx*の精度管理

L1 (Low / Bas / Bajo / Niedrig / Basso / 低濃度)

LOT 60190804  04/12/26

L2 (Normal / Normale / 正常濃度)

LOT 60190805  04/12/26

| Parameter Paramètre Parámetro Parameter Parametro 項目 | Target Valeur cible Objetivo Ergebnis Valore ideale 目標値 | Expected Range Intervalle attendu Rango esperado Referenzbereich Intervallo atteso 基準範囲 |
|---|--|--|
| RBC/GR (10 ⁹ /μL) | 2.21 | 2.06 - 2.36 |
| HCT (%) | 14.4 | 11.4 - 17.4 |
| HGB (g/dL) | 4.7 | 4.4 - 5.0 |
| MCV/VGM (fL) | 65.1 | 57.6 - 72.6 |
| MCH/TCMH (pg) | 21.3 | 17.5 - 25.1 |
| MCHC/CCMH (g/dL) | 32.6 | 21.4 - 43.8 |
| PLT (K/μL) | 53 | 23 - 83 |
| RDW-SD (fL) | 45.8 | 36.8 - 54.8 |
| RDW-CV (%) | 20.7 | 17.2 - 24.2 |
| PDW (fL) | 10.1 | 6.1 - 14.1 |
| PCT (%) | 0.05 | 0.03 - 0.07 |
| MPV (fL) | 8.6 | 5.6 - 11.6 |
| P-LCR (%) | 12.6 | 5.1 - 20.1 |
| WBC/GB (K/μL) | 2.97 | 2.57 - 3.37 |
| NEUT# (K/μL) | 1.43 | 1.02 - 1.84 |
| LYMPH# (K/μL) | 0.96 | 0.56 - 1.36 |
| MONO# (K ⁹ /μL) | 0.28 | 0.00 - 0.58 |
| EO# (K/μL) | 0.30 | 0.10 - 0.50 |
| RBC-O (10 ⁹ /μL) | 2.20 | 1.80 - 2.60 |
| PLT-O (K/μL) | 41 | 11 - 71 |
| RETIC (K/μL) | 147.3 | 107.3 - 187.3 |
| % RETIC (%) | 6.67 | 4.97 - 8.37 |

| Parameter Paramètre Parámetro Parameter Parametro 項目 | Target Valeur cible Objetivo Ergebnis Valore ideale 目標値 | Expected Range Intervalle attendu Rango esperado Referenzbereich Intervallo atteso 基準範囲 |
|---|--|--|
| RBC/GR (10 ⁹ /μL) | 4.35 | 4.08 - 4.62 |
| HCT (%) | 34.8 | 31.3 - 38.3 |
| HGB (g/dL) | 11.5 | 11.0 - 12.0 |
| MCV/VGM (fL) | 79.9 | 71.9 - 87.9 |
| MCH/TCMH (pg) | 26.4 | 22.7 - 30.1 |
| MCHC/CCMH (g/dL) | 33.0 | 26.9 - 39.1 |
| PLT (K/μL) | 207 | 162 - 252 |
| RDW-SD (fL) | 49.8 | 42.8 - 56.8 |
| RDW-CV (%) | 18.7 | 15.7 - 21.7 |
| PDW (fL) | 10.0 | 7.0 - 13.0 |
| PCT (%) | 0.19 | 0.11 - 0.27 |
| MPV (fL) | 9.3 | 6.3 - 12.3 |
| P-LCR (%) | 14.6 | 8.7 - 20.5 |
| WBC/GB (K/μL) | 7.39 | 6.64 - 8.14 |
| NEUT# (K/μL) | 3.70 | 2.64 - 4.76 |
| LYMPH# (K/μL) | 2.17 | 1.55 - 2.79 |
| MONO# (K ⁹ /μL) | 0.77 | 0.01 - 1.53 |
| EO# (K/μL) | 0.75 | 0.39 - 1.11 |
| RBC-O (10 ⁹ /μL) | 4.22 | 3.52 - 4.92 |
| PLT-O (K/μL) | 171 | 121 - 221 |
| RETIC (K/μL) | 118.2 | 88.2 - 148.2 |
| % RETIC (%) | 2.72 | 1.92 - 3.52 |

Patent information: idexx.com/patents.

© 2013 IDEXX Laboratories, Inc. All rights reserved. • 06-0002107-00
 *e-CHECK is a trademark of Sysmex Corporation. ProCyte Dx is a trademark or registered trademark of IDEXX Laboratories, Inc. or its affiliates in the United States and/or other countries.

