# e-CHECK* (XS) Quality Control for the **ProCyte Dx*** Analyzer
# e-CHECK* (XS) contrôle qualité pour l'analyseur **ProCyte Dx***
# e-CHECK* (XS) Control de calidad para el analizador **ProCyte Dx***
# e-CHECK* (XS) Qualitätskontrolle für das **ProCyte Dx*** Analysegerät
# e-CHECK* (XS) Controllo qualita' per l'analizzatore **ProCyte Dx***
# e-CHECK* (XS) IDEXX**プロサイトDx***の精度管理

**L1 (Low / Bas / Bajo / Niedrig / Basso / 低濃度)**

LOT 62990804 01/17/27

| Parameter<br>Paramètre<br>Parámetro<br>Parameter<br>Parametro<br>項目 | Target<br>Valeur cible<br>Objetivo<br>Ergebnis<br>Valore ideale<br>目標値 | Expected Range<br>Intervalle attendu<br>Rango esperado<br>Referenzbereich<br>Intervallo atteso<br>基準範囲 |
|---|---|---|
| RBC/GR ($10^6/\mu L$) | 2.28 | 2.13 - 2.43 |
| HCT (%) | 15.7 | 12.7 - 18.7 |
| HGB (g/dL) | 5.0 | 4.7 - 5.3 |
| MCV/VGM (fL) | 68.9 | 61.4 - 76.4 |
| MCH/TCMH (pg) | 21.9 | 18.1 - 25.7 |
| MCHC/CCMH (g/dL) | 31.8 | 20.6 - 43.0 |
| PLT (K/$\mu$L) | 54 | 24 - 84 |
| RDW-SD (fL) | 44.4 | 35.4 - 53.4 |
| RDW-CV (%) | 18.8 | 15.3 - 22.3 |
| PDW (fL) | 8.5 | 4.5 - 12.5 |
| PCT (%) | 0.04 | 0.02 - 0.06 |
| MPV (fL) | 8.0 | 5.0 - 11.0 |
| P-LCR (%) | 8.5 | 1.0 - 16.0 |
| WBC/GB (K/$\mu$L) | 2.87 | 2.47 - 3.27 |
| NEUT# (K/$\mu$L) | 1.24 | 0.83 - 1.65 |
| LYMPH# (K/$\mu$L) | 1.07 | 0.67 - 1.47 |
| MONO# ($K^9/\mu L$) | 0.28 | 0.00 - 0.58 |
| EO# (K/$\mu$L) | 0.28 | 0.08 - 0.48 |
| RBC-O ($10^6/\mu L$) | 2.31 | 1.91 - 2.71 |
| PLT-O (K/$\mu$L) | 42 | 12 - 72 |
| RETIC (K/$\mu$L) | 149.7 | 109.7 - 189.7 |
| % RETIC (%) | 6.57 | 4.87 - 8.27 |

**L2 (Normal / Normale / 正常濃度)**

LOT 62990805 01/17/27

| Parameter<br>Paramètre<br>Parámetro<br>Parameter<br>Parametro<br>項目 | Target<br>Valeur cible<br>Objetivo<br>Ergebnis<br>Valore ideale<br>目標値 | Expected Range<br>Intervalle attendu<br>Rango esperado<br>Referenzbereich<br>Intervallo atteso<br>基準範囲 |
|---|---|---|
| RBC/GR ($10^6/\mu L$) | 4.32 | 4.05 - 4.59 |
| HCT (%) | 35.3 | 31.8 - 38.8 |
| HGB (g/dL) | 11.8 | 11.3 - 12.3 |
| MCV/VGM (fL) | 81.8 | 73.8 - 89.8 |
| MCH/TCMH (pg) | 27.3 | 23.6 - 31.0 |
| MCHC/CCMH (g/dL) | 33.4 | 27.3 - 39.5 |
| PLT (K/$\mu$L) | 211 | 166 - 256 |
| RDW-SD (fL) | 46.4 | 39.4 - 53.4 |
| RDW-CV (%) | 16.7 | 13.7 - 19.7 |
| PDW (fL) | 9.1 | 6.1 - 12.1 |
| PCT (%) | 0.18 | 0.10 - 0.26 |
| MPV (fL) | 8.7 | 5.7 - 11.7 |
| P-LCR (%) | 10.7 | 4.8 - 16.6 |
| WBC/GB (K/$\mu$L) | 7.12 | 6.37 - 7.87 |
| NEUT# (K/$\mu$L) | 3.48 | 2.42 - 4.54 |
| LYMPH# (K/$\mu$L) | 2.16 | 1.54 - 2.78 |
| MONO# ($K^9/\mu L$) | 0.74 | 0.00 - 1.50 |
| EO# (K/$\mu$L) | 0.74 | 0.38 - 1.10 |
| RBC-O ($10^6/\mu L$) | 4.22 | 3.52 - 4.92 |
| PLT-O (K/$\mu$L) | 168 | 118 - 218 |
| RETIC (K/$\mu$L) | 111.8 | 81.8 - 141.8 |
| % RETIC (%) | 2.59 | 1.79 - 3.39 |

Patent information: idexx.com/patents.

© 2013 IDEXX Laboratories, Inc. All rights reserved. • 06-0002107-00
*e-CHECK is a trademark of Sysmex Corporation. ProCyte Dx is a trademark or registered trademark of IDEXX Laboratories, Inc. or its affiliates in the United States and/or other countries.

IDEXX BioResearch